

MMT-8007A - DIGITAL ALARM-THERMOMETER WITH GLYCOL PROBE

SPECIFICATION

1.Measuring temperature range:

- External Glycol probe Temperature Range: -50~+70 °C/-58~+158 °F
- Internal sensor Temperature Range: 0~+50 °C/+32~+122 °F

2.Temperature accuracy: ±1°C (±2°F)

3.Temperature display resolution: 0.1°C (0.1°F)

4.MAX/MIN temperature memory records

5.Operating voltage:1.5V, 1 PCS AAA battery

6.Total Dimension:110mm*70mm*19.5mm

7.Net weight: pprox.82g

FEATURES

1.LCD screen shows indoor & outdoor temperature.

2.Preset alert temperature.

3.Stand for shelf or hanging hole at back.

4.Max/Min temperature memory records.

BUTTONS&OPERATIONS

1.MIN/MAX

(1) Press MIN/MAX key to display the maximum indoor temperature.

(2) Press MIN/MAX key again to display the minimum indoor temperature

2.RESET

(1) Reset max and min to current temperature.

(2) Reset alert temperature range to -1°C~4°C

3.ALERT SET

(1) Turn on/off ALERT

(2) Holding 1 second for setting alert temperature range

4.°C/°F & ADJ

(1) Adding one deg C press once or auto advanced (-50~70) when hold the key under ALERT SET mode.

(2) Select °C or °F Display



1295 Morningside Avenue, Unit 16-18
Scarborough, ON M1B 4Z4 Canada
Phone: 416-261-4865 Fax: 416-261-7879
www.scigiene.com

ALERT SET

ALERT
SET

HOLDING 1 SECOND FOR SETTING ALERT
TEMPERATURE TANGE

4.0 IN
0.0 OUT

C/F
ADJ

ADDING ONE DEG C PRESS ONCE OR AUTO
ADVANCED (-50~+70) WHEN HOLDING THE KEY

4.0 IN
0.0 OUT

C/F
ADJ

HOLDING THE KEY

4.0 IN
24.0 OUT

ALERT
SET

4.0 IN
24.0 OUT

C/F
ADJ

HOLDING THE KEY

26.0 IN
24.0 OUT

ALERT
SET

Buzzer on 5 seconds and off 55 seconds
when the Outdoor Temp.
between(+24degC~ +26degC.)

21.8 IN
24.0 OUT

ALERT
SET

HOLDING 1 SECOND FOR SETTING ALERT

26.0 IN
24.0 OUT

C/F
ADJ

HOLDING THE KEY

26.0 IN
28.0 OUT

ALERT
SET

ALERT
SET

Buzzer on 5seconds and off
55 seconds when the Outdoor
Temp.<26degC or> 28degC

21.8 IN
24.0 OUT

ALERT
SET

TURN OFF ALERT

ALE

OFF

21.8 IN
24.0 OUT

ALERT
SET

TURN ON ALERT

ALE

ON

21.8 IN
24.0 OUT